

<b>Computation Constants</b>		Comp Constants for Edwards CO Set+ Closed Injectate System	Comp Constants for bath probes inserted in room temp fluid	
PA Catheter Model		Room temp injectate (18-25°C)	Room temp injectate (19-22/23-25)	
VIP+	(834HF75)	0.595	0.582	0.607
paceport	(991HF8)	0.607	na	0.616
Heparin free	(139F75)	0.595	0.582	0.594
Latex free	(C146F7)	CO Set <b>not</b> latex free	0.579	0.597
TD Pacing Swan	(D200HF7)	0.608	0.578	0.595
VIP	(831HF75)	0.592	0.588	0.612
CCO	(777HF8)	0.602	0.585	0.600
Arrow 7.5 LEAVE IN SWAN		0.586	NA	
Arrow 7.0 LEAVE IN SWAN		0.601	NA	
Arrow 6.0 LEAVE IN SWAN		0.621	NA	